

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 04 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130303</u> ✓
Name		<u>JEAN ACHEE</u>	ID# <u>1470</u> Date <u>3/1/2010</u> ✓
A	Agency	<u>WASILLA POLICE DEPT</u>	Phone # <u>352-5401</u>
	Instrument Location	<u>WASILLA POLICE DEPT</u>	
B	Alco S/N	<u>G30491</u> ✓	Target Value <u>.083</u> ✓ High Pressure <u>1000</u>
	Alco Test Value Average	<u>.081</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco
Signature		<u>Jean Achee</u>	
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program


Subscribed and sworn before me this 28<sup>th</sup> day of July, 2010.

(Notary Seal Stamp)

Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office  
TARGET VALUE NOT PROGRAMMED.  
BREATH PERSONNEL NEEDED TO  
AUTHENTICATE V&C FOR COURT.

BT9 03/00

NJB 7/28/10



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

MARCH 01, 2010

OPERATOR'S NAME:

ACHEE, JEAN P

OPERATOR'S NUMBER: 1470

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567890

DEPT/AGENCY: WAS1

CASE/REPORT: 10-0301

TEST TYPE: V

ALCO TARGET VALUE: .083

ALCO S/N: G30491

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 28.86 in  
ALCO TARGET .076  
BLANK TEST .000  
INTERNAL STANDARD VERIFIED  
ALCO TV 28.86 in .081  
BLANK TEST .000  
SUBJECT SAMPLE .000  
BLANK TEST .000  
ALCO TV 28.86 in .079  
BLANK TEST .000

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

MARCH 01, 2010

TIME 15:07

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 28.86 in  
FLOW DETECTOR: OKAY  
15:11  
15:12 PUMP  
15:12 HIGH SPEED: OKAY  
15:12  
15:13 DETECTOR: OKAY  
15:13  
15:14 FILTERS: OKAY  
15:14  
15:15 QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDE  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~